



OFFICE OF THE PROCUREMENT OFFICER
ABDUL WALI KHAN UNIVERSITY MARDAN

Phone: 0937-920868

No/AWKUM/Proc/2022/

Date: 14/07/2022

ADDENDUM # 1

PRE-BID MEETING MINUTES FOR PURCHASE OF LAB EQUIPMENTS FOR ESTABLISHMENT UNDER HEC FUNDED PROJECT STRENGTHING OF ABDUL WALI KHAN UNIVERSITY MARDAN, HELD ON JULY 14, 2022

The following Members participated in the meeting;

- | | |
|---|-----------|
| 1. Prof. Dr. Saeed Islam, Dean of Physical & Numerical Sciences | Convener |
| 2. Mr. Sayim Sohail, Registrar | Member |
| 3. Dr. Muhammad Idrees, Director Administration | Member |
| 4. Dr. Shaista Irshad Khan, Treasurer | Member |
| 5. Dr. Sajid Ali, Associate Professor | Member |
| 6. Mr. Kamran Ullah, Procurement Officer | Secretary |

The following Firms representative participated in the meeting;

1. M/S Kasbn International
2. M/S Meditec Instrument
3. M/S Raheem Scientific
4. M/S Leo Science
5. M/S AMS Pvt Ltd.
6. M/S Pakistan Microbiological Associates
7. M/S Snibe Diagnostics
8. M/S Friend Traders Peshawar

The queries and response of the bidders and university are as follows;

S/No.	Queries Raised by prospective bidders	Response
1.	<p>1) The ELISA Reader and Washer should be separated as these are two different equipment. Please also note that most of the good manufacturers of Readers do not manufacture Washers and it is therefore unpractical to quote equipment from two different manufacturers under a same heading.</p> <p>2) The specifications of ELISA washer are not given and should be provided after separation of these two equipment.</p> <p>3) The 6' touch screen for ELISA Reader should not be included as a minimum requirement because now a days latest and high-tech equipment are being made to be controlled through computers and smart phones. Therefore, please omit the 6' display as the minimum requirement.</p>	<p>1) Washer and Elisa reader, we need combine system from the same firm/manufacturer as we will face problems in future by purchasing it from different companies and many are there that they can supply.</p> <p>2) Specifications of washer, anyone who can bid Elisa can quote washer accordingly.</p> <p>3) The 6 inch screen is minimal requirement. Anyone can quite above than this or with advance technology screen.</p>

2.	Bio chemistry analyzer there is a mention of more than 2 channels on ISE module. As the required Equipment is a medium end routine chemistry analyzer it is not necessary to have an ISE module with 2 channels. Only one equipment Roche C111 with such low sample thought put has the options which leads to favoring of a specified equipment in the tender. Kindly revise the specs to "single channel equipment" to enable others to participate & offer a fair & healthy competition.	Specifications of Bio Chemistry analyzer remains the same as per Tender Documents
3.	As per pre-bid meeting it was said there will be a tender requirement to attach PO's in technical bid that shall have prices not hidden, so we can see that your business is above 5 million whereas there is no such tender requirement anywhere across Pakistan in any of the major public sector hospitals. To estimate or have an idea of the total business hospital can see the business volume from our annual return & audit reports. Further in annual returns, the sales figures are mentioned. The overall sales of the company can be easily estimated from these documents and does not require any purchase orders with visible prices in the technical bid as this will automatically create a financial impact during technical evaluation and bidders will not be evaluated only on the technical basis.	To verify the amount of the PO, it is necessary to have the amount written in the PO. So, all the bidders have to submit PO's with the amount mentioned. Otherwise no marks will be granted.
4.	Product certification Such as FDA,MHLW and JAIS are required, However, CE certification may be added as well.	Any one of the certification is mandatory. Additional certification will carry marks
5.	-86 C Ultra low temperature Freezer capacity is not mentioned.	520 liters to 560 Liters capacity.
6.	High pressure Steam Sterilizer capacity is not mentioned.	50 Liters capacity.